



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

AMERICAL, INC.
100 Ferncroft Rd., Unit #205
Danvers, MA 01923
Michael Papageorgiou Phone: 978 646 9165

CALIBRATION

Valid To: August 31, 2020

Certificate Number: 2714.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Dimensional

Parameter/Equipment	Range	CMC ² (±)	Comments
Micrometers – Outside	Up to 1 in	67 µin	Gage blocks and master ball
Calipers – Dial and Digital	Up to 6 in	510 µin	Gage blocks and master ring gage

¹ This laboratory offers commercial calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.



Accredited Laboratory

A2LA has accredited

AMERICAL, INC.

Danvers, MA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – *Specific Requirements: Calibration Laboratory Accreditation Program*. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 13th day of September 2018.

A handwritten signature in blue ink, written over a horizontal line.

President and CEO
For the Accreditation Council
Certificate Number 2714.01
Valid to August 31, 2020

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.